visocolor[®] HE

Nitrite

High sensitivity test kit for the determination in the range of $0.005-0.10~mg/L~NO_2^-$

Method:

Azo dye formed by sulfanilamide and N-(1-naphthyl) ethylene diamine Contents of test kit (*refill pack):

sufficient for 150 tests

100 mL NO₂-1

- 23 g NO₂-2*
 - 1 black measuring spoon 85 mm*
 - 1 plastic beaker for sampling
 - 2 round glass tubes with screw caps
 - 1 comparator block
 - 1 color comparison disc Nitrite

Hazard warning:

Information regarding safety can be found on the box' label and in the safety data sheet. You can download the SDS from *www.mn-net.com/SDS*.

Procedure:

- 1. Insert color comparison disc (see illustration).
- Open both round glass tubes, rinse several times with the water sample and fill up to the mark with the sample.
- 3. Add 12 drops NO₂-1 to the right glass tube, close and mix.
- Add 1 level measuring spoon NO₂-2 to the right glass tube, close and mix. Wait 10 min.
- 5. Reading: Turn color disc until both colors match by transmitted light from above. Read test results from the mark on the front side of the comparator (see illustration). Intermediate values can be estimated.
- 6. After use clean both round glass tubes thoroughly and close.

mg/L NO ₂ -	mg/L NO ₂ -N	mmol/m ³
0.005	0.002	0.11
0.010	0.003	0.22
0.015	0.005	0.33
0.02	0.006	0.43
0.03	0.009	0.65
0.04	0.012	0.87
0.06	0.018	1.30
0.08	0.024	1.74
0.10	0.030	2.17

The method can be applied also for the analysis of sea water.

Disposing of the samples:

Information regarding disposal can be found in the safety data sheet. You can download the SDS from *www.mn-net.com/SDS*.

Interferences:

Chromium(VI) and iron(III) ions > 3 mg/L simulate high nitrite concentrations.

Chlorine interferes in minute concentrations.

The reaction is temperature dependent. Below 20 $^\circ \rm C$ results can be too low, above 25 $^\circ \rm C$ too high.

