**REF 985098** 

en

Test 0-98 03.23

NANOCOLOR® Aluminum 07

### Method:

Photometric determination with eriochrome cyanine R

Range: 0.02-0.70 mg/L Al<sup>3+</sup>

Wavelength (HW = 5 - 12 nm): **540 nm** 

Reaction time: 5 min (300 s)
Reaction temperature: 20-25 °C

# Contents of reagent set:

19 test tubes Aluminum 07 with NANOFIX

1 test tube with 11 mL Aluminum 07 R2

1 test tube with 11 mL Aluminum 07 R3

1 test tube with 11 mL Aluminum 07 R4

1 test tube with blank value "NULL"

## **Hazard warning:**

This test does not contain any harmful substances which must be specially labelled as hazardous.

## **Preliminary tests:**

If the order of magnitude of the concentration in a sample is not known, a preliminary test with QUANTOFIX® Aluminum (5–500 mg/L Al $^{3+}$ , REF 91307) rapidly gives this information. From the order of magnitude the required dilution can be calculated and prepared directly.

### Interferences:

Fluoride interferes.

Turbid samples are to be filtered (membrane filter 0.45  $\mu$ m, REF 91650). The total Aluminum can be determined with NANOCOLOR® NanOx Metal (REF 918978) and microwave decomposition.

The following ions will not interfere:

≤ 100 mg/L SiO<sub>2</sub>

 $\leq$  10 mg/L Cu<sup>2+</sup>, Fe<sup>2/3+</sup>, Mn<sup>2+</sup>, Ni<sup>2+</sup>, Zn<sup>2+</sup>

 $\leq$  5 mg/L Cr<sup>3+</sup>, Cd<sup>2+</sup>

≤ 1 mg/L Co<sup>2+</sup>

The method can be applied also for the analysis of sea water.

#### Procedure:

Requisite accessories: piston pipette with tips

Open test tube, add

**500**  $\mu$ L (= 0.5 mL) R2,

**500**  $\mu$ L (= 0.5 mL) R3 and

4.0 mL sample solution (the pH value of the sample must be between pH 3 and 6), close and mix.

Open test tube again, add

**500** μL (= 0.5 mL) R4, close and mix.

Clean outside of test tube and measure after 5 min.

Adjust photometer to zero by using blank value "NULL".

### Measurement:

For NANOCOLOR® photometers and PF-12 see manual, test 0-98.

### Measurement when samples are colored or turbid:

For all NANOCOLOR® photometers see manual, use key for correction value.

### Photometers of other manufacturers:

For other photometers check whether measurement of round glass tubes is possible. Verify factor for each type of instrument by measuring standard solutions. The factor depends strongly on the wavelength.

## Analytical quality control:

NANOCONTROL Multistandard Drinking water (REF 925018)